

CONEX-AGP

*Agilis-P Controller
with Encoder Feedback*



Newport®
Experience | Solutions

**Introduction,
Safety & Service**

Table of Contents

Preface.....iv

Documentation Organization.....iv

Confidentiality & Proprietary Rightsiv

Sales, Tech Support & Servicev

EU Declaration of Conformityvi

1.0 Introduction 1

1.1 Definitions and Symbols.....1

1.1.1 General Warning or Caution.....1

1.1.2 Electric Shock.....1

1.1.3 European Union CE Mark1

1.2 Warnings and Cautions2

1.3 General Warnings and Cautions2

2.0 Warranty 3

3.0 Maintenance and Service 4

3.1 Enclosure Cleaning4

3.2 Obtaining Service4

Service Form 5

Preface

Documentation Organization

CONEX-AGP documentation set includes 2 parts:

- **Introduction, Safety and Service** (this document).
- **Controller Manual** describing available commands.

NOTE

During operation, always refer to this document for safety information.

Confidentiality & Proprietary Rights

Reservation of Title

The Newport Programs and all materials furnished or produced in connection with them ("Related Materials") contain trade secrets of Newport and are for use only in the manner expressly permitted. Newport claims and reserves all rights and benefits afforded under law in the Programs provided by Newport Corporation.

Newport shall retain full ownership of Intellectual Property Rights in and to all development, process, align or assembly technologies developed and other derivative work that may be developed by Newport. Customer shall not challenge, or cause any third party to challenge, the rights of Newport.

Preservation of Secrecy and Confidentiality and Restrictions to Access

Customer shall protect the Newport Programs and Related Materials as trade secrets of Newport, and shall devote its best efforts to ensure that all its personnel protect the Newport Programs as trade secrets of Newport Corporation. Customer shall not at any time disclose Newport's trade secrets to any other person, firm, organization, or employee that does not need (consistent with Customer's right of use hereunder) to obtain access to the Newport Programs and Related Materials. These restrictions shall not apply to information (1) generally known to the public or obtainable from public sources; (2) readily apparent from the keyboard operations, visual display, or output reports of the Programs; (3) previously in the possession of Customer or subsequently developed or acquired without reliance on the Newport Programs; or (4) approved by Newport for release without restriction.

Sales, Tech Support & Service

North America & Asia

Newport Corporation
1791 Deere Ave.
Irvine, CA 92606, USA

Sales

Tel.: (800) 222-6440
e-mail: sales@newport.com

Technical Support

Tel.: (800) 222-6440
e-mail: tech@newport.com

Service, RMAs & Returns

Tel.: (800) 222-6440
e-mail: rma.service@newport.com

Europe

MICRO-CONTROLE Spectra-Physics S.A.S
1, rue Jules Guesde – Bât. B
ZI Bois de l'Épine – BP189
91006 Evry Cedex
France

Sales France

Tel.: +33 (0)1.60.91.68.68
e-mail: france@newport-fr.com

Sales Germany

Tel.: +49 (0) 61 51 / 708 – 0
e-mail: germany@newport.com

Sales UK

Tel.: +44 (0)1635.521757
e-mail: uk@newport.com



Technical Support

e-mail: tech_europe@newport.com

Service & Returns

Tel.: +33 (0)2.38.40.51.55

EU Declaration of Conformity

CONEX-AGP	 Newport Experience Solutions
EC Declaration of Conformity following Annex II-1A of Directive 2006/42/EC on machinery	
The manufacturer: MICRO-CONTROLE Spectra-Physics, 1 rue Jules Guesde ZI Bois de l'Épine - BP189 F-91006 Evry FRANCE	
Hereby declares that the machinery: <ul style="list-style-type: none">• Description: "CONEX-AGP"• Function: Agilis-P Controller + Mechanical• Models: AG-LS25-27P ; AG-PR100P	
– the technical file of which was compiled by: Mr Dominique DEVIDAL, Quality Director, MICRO-CONTROLE Spectra-Physics, Zone Industrielle - B.P.29 F-45340 Beaune La Rolande France	
– complies with all the relevant provisions of the Directive 2006/42/EC on machinery. – complies with all the relevant provisions of the Directive 2004/108/EC relating to electro-magnetic compatibility.	
– was designed and built in accordance with the following harmonised standards: <ul style="list-style-type: none">• NF EN 61326-1:2006 « Electrical equipment for measurement, control and laboratory use – EMC requirements – Part 1: General requirements »• NF EN 55011:2007 Class A	
– was designed and built in accordance with the following other standards: <ul style="list-style-type: none">• NF EN 61000-4-2:2009• NF EN 61000-4-3:2006 + A2:2008• NF EN 61000-4-4:2005• NF EN 61000-4-6:2009	
ORIGINAL DECLARATION	
Done in Beaune La Rolande on 30 July 2010 Dominique DEVIDAL Quality Director	
DC1-EN rev:A	

CONEX-AGP

Agilis-P controller with encoder feedback

1.0 Introduction

1.1 Definitions and Symbols

The following terms and symbols are used in this documentation and also appear on the CONEX-CC Controller/Driver where safety-related issues occur.

1.1.1 General Warning or Caution



Figure 1: General Warning or Caution Symbol.

The Exclamation Symbol in Figure 1 may appear in Warning and Caution tables in this document. This symbol designates an area where personal injury or damage to the equipment is possible.

1.1.2 Electric Shock



Figure 2: Electrical Shock Symbol.

The Electrical Shock Symbol in Figure 2 may appear on labels affixed to the CONEX-CC Controller/Driver. This symbol indicates a hazard arising from dangerous voltage. Any mishandling could result in irreparable damage to the equipment, in personal injury, or death.

1.1.3 European Union CE Mark



Figure 3: CE Mark.

The presence of the CE Mark on Newport Corporation equipment means that it has been designed, tested and certified as complying with all applicable European Union (CE) regulations and recommendations.

1.2 Warnings and Cautions

The following are definitions of the Warnings, Cautions and Notes that may be used in this manual to call attention to important information regarding personal safety, safety and preservation of the equipment, or important tips.



WARNING

Situation has the potential to cause bodily harm or death.



CAUTION

Situation has the potential to cause damage to property or equipment.

NOTE

Additional information the user or operator should consider.

1.3 General Warnings and Cautions

The following general safety precautions must be observed during all phases of operation of this equipment.

Failure to comply with these precautions or with specific warnings elsewhere in this manual violates safety standards of design, manufacture, and intended use of the equipment.

- Heed all warnings on the unit and in the operating instructions.
- To prevent damage to the equipment, read the instructions in this manual.
- Route power cords and cables where they are not likely to be damaged.
- Keep air vents free of dirt and dust.
- Keep liquids away from unit.
- Do not expose equipment to excessive moisture (>85% humidity)
- Do not operate this equipment in an explosive atmosphere.
- Disconnect power before cleaning the CONEX-AGP unit. Do not use liquid or aerosol cleaners.
- Do not open the CONEX-AGP controller. There are no user-serviceable parts inside.
- Return equipment to Newport Corporation for service and repair.
- Follow precautions for static-sensitive devices when handling electronic circuits.

2.0 Warranty

Newport Corporation warrants that this product will be free from defects in material and workmanship and will comply with Newport's published specifications at the time of sale for a period of one year from date of shipment. If found to be defective during the warranty period, the product will either be repaired or replaced at Newport's option.

To exercise this warranty, write or call your local Newport office or representative, or contact Newport headquarters in Irvine, California. You will be given prompt assistance and return instructions. Send the product, freight prepaid, to the indicated service facility. Repairs will be made and the instrument returned freight prepaid. Repaired products are warranted for the remainder of the original warranty period or 90 days, whichever occurs last.

Limitation of Warranty

The above warranties do not apply to products which have been repaired or modified without Newport's written approval, or products subjected to unusual physical, thermal or electrical stress, improper installation, misuse, abuse, accident or negligence in use, storage, transportation or handling.

THIS WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESSED OR IMPLIED, INCLUDING ANY IMPLIED WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. NEWPORT CORPORATION SHALL NOT BE LIABLE FOR ANY INDIRECT, SPECIAL, OR CONSEQUENTIAL DAMAGES RESULTING FROM THE PURCHASE OR USE OF ITS PRODUCTS.

First printing 2010.

Copyright 2011 by Newport Corporation, Irvine, CA. All rights reserved. No part of this manual may be reproduced or copied without the prior written approval of Newport Corporation. This manual is provided for information only, and product specifications are subject to change without notice. Any change will be reflected in future printings.

3.0 Maintenance and Service

3.1 Enclosure Cleaning

The CONEX-AGP Controller/Driver should only be cleaned with a lightly dampened cloth or sponge with a soapy water solution. Do not use an acetone or alcohol solution; this will damage the finish of the enclosure.

3.2 Obtaining Service

The user should not attempt any maintenance or service of the CONEX-AGP Controller/Driver and its accessories beyond the procedures outlined in this manual.

Any problem that cannot be resolved should be referred to Newport Corporation. When calling Newport regarding a problem, please provide the Tech Support representative with the following information:

- Your contact information.
- System serial number or original order number.
- Description of problem.
- Environment in which the system is used.
- State of the system before the problem.
- Frequency and repeatability of problem.
- Can the product continue to operate with this problem?
- Can you identify anything that may have caused the problem?

Complete a copy of the Service Form as represented on the next page and include it with your shipment.

Any CONEX-AGP controller/driver being returned to Newport must have been assigned an RMA number by Newport. Assignment of the RMA requires the item serial number.

CONEX-AGP Controller/Driver being returned under an RMA must be securely packaged for shipment. If possible, reuse the original factory packaging.

Your Local Representative

Fax: _____

Return authorization #: _____

Date: _____

Phone Number: _____


Fax Number: _____

Serial #: _____

Serial #: _____

Reasons of return of goods (please list any specific problems): _____

This image shows a single sheet of white paper with horizontal ruling lines. The lines are evenly spaced and run across the width of the page. There are no margins, text, or other markings on the paper.



Newport®
Experience | Solutions

Visit Newport Online at:
www.newport.com

North America & Asia

Newport Corporation
1791 Deere Ave.
Irvine, CA 92606, USA

Sales

Tel.: (800) 222-6440
e-mail: sales@newport.com

Technical Support

Tel.: (800) 222-6440
e-mail: tech@newport.com

Service, RMAs & Returns

Tel.: (800) 222-6440
e-mail: rma.service@newport.com

Europe

MICRO-CONTROLE Spectra-Physics S.A.S
1, rue Jules Guesde – Bât. B
ZI Bois de l'Épine – BP189
91006 Evry Cedex
France

Sales

Tel.: +33 (0)1.60.91.68.68
e-mail: france@newport-fr.com

Technical Support

e-mail: tech_europe@newport.com

Service & Returns

Tel.: +33 (0)2.38.40.51.55